

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 23 2009

DataMaster cdm S/N <u>130179</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>James G. Truesdell</u>	ID# <u>2856</u> Date <u>11/22/09</u>
A Agency <u>Alaska State Troopers</u>	Phone # <u>262-4453</u>
Instrument Location <u>Soldotna AST Post</u>	
Alco S/N <u>X124630</u> Target Value <u>.079</u> High Pressure <u>500</u>	
B Alco Test Value Average <u>.077</u>	<u>.077</u>
1 st Alco	2 nd Alco
Signature <u>James G. Truesdell</u>	
(OVER)	

BMB
12/31/09


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 31st day of Dec, 2009.

Carolyn M. Noland
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

NOVEMBER 22, 2009

OPERATOR'S NAME: BMB
TRUESDELL/JAMES/G
OPERATOR'S NUMBER: 2856
SUBJECT'S LAST NAME: 12/31/09
VERIFICATION
SUBJECT'S FIRST NAME/MI :
OF/CAL
O.L. #: 1234567
DEPT/AGENCY: SLD0
CASE/REPORT: 00000000
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X124630

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.46 in
ALCO TARGET .077 09:53
BLANK TEST .000 09:54
INTERNAL STANDARD VERIFIED 09:54
ALCO TV 29.46 in .077 09:55
BLANK TEST .000 09:56
SUBJECT SAMPLE .000 09:56
BLANK TEST .000 09:57
ALCO TV 29.47 in .077 09:57
BLANK TEST .000 09:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130179

NOVEMBER 22, 2009
TIME 09:59

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.47 in
FLAM DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`'abcde fghijklmno
pqrstuvwxyz{|}~!@